



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/524,607	07/20/2005	Satoru Hashimoto	J2000-7001US	1117
37462 7590 02/28/2007 LOWRIE, LANDO & ANASTASI RIVERFRONT OFFICE ONE MAIN STREET, ELEVENTH FLOOR CAMBRIDGE, MA 02142			EXAMINER BOUCHELLE, LAURA A	
			ART UNIT 3763	PAPER NUMBER
SHORTENED STATUTORY PERIOD OF RESPONSE			MAIL DATE	DELIVERY MODE
3 MONTHS			02/28/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

If NO period for reply is specified above, the maximum statutory period will apply and will expire 6 MONTHS from the mailing date of this communication.

Office Action Summary

Application No.

10/524,607

Applicant(s)

HASHIMOTO ET AL.

Examiner

Laura A. Bouchelle

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 20 November 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-3 and 5-17 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-3, 5-9 and 12-17 is/are rejected.
- 7) ☒ Claim(s) 10 and 11 is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

2. Claims 1-3, 5, 6, 12-15 are rejected under 35 U.S.C. 102(e) as being anticipated by Larrain et al (US 2002/0002350). Larrain discloses a pressure valve comprising a first fluid path 12, a second fluid path 18, a housing 10, a valve member having a body 20 and a hemispherical projection portion 24. The body of the valve has a tubular member that is inserted into the first fluid path and a communication portion having a concave shape. The valve member comprises a resilient silicon rubber material (Page 2, paragraph 0033). The valve member is opened by a pressure below 0.2 Kgf/cm² (Abstract). The valve member is placed under tension in the lengthwise direction (Page 3, paragraph 0036).

Claim Rejections - 35 USC § 103

3. The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Art Unit: 3763

4. Claims 7, 8, 9 are rejected under 35 U.S.C. 103(a) as being unpatentable over Larrain in view of Choksi (US 5727594). Claim 7 differs from Larrain in calling for the second fluid path to have a conical shape. Claim 8 differ in calling for the first and second housing members to be attached without the use of adhesives. Choksi teaches a check valve having a first housing member and a second housing member, the second housing member having a conical shape that allows the valve to be attached to a standard adapter, and the first and second housing members attached without the use of adhesive to reduce manufacturing costs (Col. 7, lines 56-61). Therefore, it would have been obvious to one of ordinary skill in the art at the time of invention to modify the device of Larrain to include a conically shaped second housing member and a first and second housing member attached without adhesive as taught by Choksi.

5. Claims 16 and 17 are rejected under 35 U.S.C. 103(a) as being unpatentable over Brown (US 6290681) in view of Larrain. Brown discloses a drug supply line having a drug supply means 84, a tubular member, a drug administering means 80, a transducer for measuring the pressure inside the tube and a pressure value display (Col. 2, lines 30-55). Claim 16 differs from Brown in calling for the valve of claim 1. Larrain teaches the valve of claim 1 as discussed above to prevent back flow of liquid (See Abstract). Therefore, it would have been obvious to one of ordinary skill in the art at the time of invention to modify the device of Brown to include the valve of Larrain to prevent backflow in the device.

Art Unit: 3763

Allowable Subject Matter

6. Claims 10, 11 are objected to as being dependent upon a rejected base claim, but would be allowable if rewritten in independent form including all of the limitations of the base claim and any intervening claims.

Response to Arguments

7. Applicant's arguments, see pages 6-8, filed 11/20/06, with respect to the rejection(s) of claim(s) 1-3,7-9,12,13 under Inoue have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made in view of Larrain.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event,

Art Unit: 3763

however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Laura A. Bouchelle whose telephone number is 571-272-2125. The examiner can normally be reached on Monday-Friday 8-4.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 517-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Laura A Bouchelle
Examiner
Art Unit 3763

